



Multi-Analyte Calibrator

Cat. No. PH 7000/7001

Lot: 20804 Exp: AUG 15

Target Value

Constituent/Method	Level 1	Level 2	Units
Albumin	3.1	6.1	g/dl
Bilirubin Direct:	2.0	4.2	mg/dL
Bilirubin Total	3.0	6.2	mg/dL
BUN (Urea Nitrogen)	34	65	mg/dL
Calcium	4.6	9.0	mg/dL
Chloride (ISE)	80	100	mEq/L
Cholesterol	175	360	mg/dL
C02	12	25	mEq/L
Creatinine	1.5	3.0	mg/dL
Glucose	85	175	mg/dL
Phosphorous	5.8	11.6	mg/dL
Magnesium	1.2	2.5	mEq/L
Total Protein	5.8	11.5	g/dL
Triglycerides	150	300	mg/dL
Uric Acid	2.9	5.2	mg/dL
Sodium (Na) (ISE)	123	158	mEq/L
Potassium (K) (ISE)	3.00	5.00	mEq/L
Iron Fe	100	200	ug/dl
HDL	75	125	mg/dL
LDL	30	50	mg/dl

- 1) The Multi Analyte Calibrator is a bovine serum based calibrator recommended for use on all chemistry analyzers
- 2) One or Two levels of calibrator are required to calibrate instruments
- 3) See value assignments as shown above for calibrator constituents and established values.
Preparation: Liquid Ready to use
- 4) Storage: 36 Months at 2 - 8 C
- 5) Open ampule Stability: 2 Days @ 2 - 8 C

8 TECH CIRCLE, NATICK, MASSACHUSETTS 01760

Tel: (508) 655 - 8310 / Fax: (508) 655 - 8273

Email: ajordan@phoenixdiagnostics.com

Web: www.phoenixdiagnostic.com